

Annexure-V

Compliance report

Note: The bidder is required to mention detailed specifications clearly in column named as Specifications from Bidder against each item as mentioned in below format. Please note that merely mentioning Yes/No in deviation column will lead to disqualification of the bidder.

(I) Name of Instrument :CORIOLLIS COMPONENT OF ACCELERATION

Sr. No.	Specification of e-Tender	Specification of Bidder/Vendor	Deviation (Yes/No)	Remarks
1	<p>CORIOLLIS COMPONENT OF ACCELERATION</p> <ul style="list-style-type: none"> ● Water Tank : Made of Stainless Steel ● Rotating Arms : 9mm/6mm orifice diameter, Length 300 mm. ● Rotameter : 250 to 2500 LPH ● Drive : Variable speed FHP motor, PMDC type ● Pump : 0.5 HP for water circulation. ● Control Panel : <ul style="list-style-type: none"> - RPM measurement : RPM Indicator with Proximity sensor. - Variac, Standard make On/Off Switch, Mains Indicator, etc. - Instruction manual : ENGLISH - consisting of experimental procedures, block diagram etc. - Rigid structure, Industrial PU Paint. 			